



Dharmsinh Desai University
R & D Center, Faculty of Technology, Nadiad.

Tender Notification

Date: 31/08/2019

Sealed tenders are invited for equipments as under,

Item 1: Syringe Pump		Quantity
Specification:	• Maximum Voltage: 24 V DC	01
	• Motor: Stepper Motor driven	
	• Minimum Resolution: 24,000/full-stroke	
	• Programmable driver and control board with serial interface	
Item 2: Distribution valve		Quantity
Specification:	• Fluid Contact Valve – Minimum 6-way (port)	01
	• Material: PCTFE / PTFE / UHMWPE	
	• psi: Minimum 30 psi	
	• Orifice: 1mm (0.04inch) diameter or less	
Item 3: Fluid Contact Syringe		Quantity
Specification:	• Capacity: 2.5 ml	01
	• Material: Borosilicate Glass / UHMWPE / PTFE	
	• psi: Minimum 30 psi	
	• Orifice: 1mm (0.04inch) diameter or less*	
Item 4: Fluid Contact Syringe		Quantity
Specification:	• Capacity: 5.0 ml	01
	• Material: Borosilicate Glass / UHMWPE / PTFE	
	• psi: Minimum 30 psi	
	• Orifice: 1mm (0.04inch) diameter or less*	

* Syringe orifice should be compatible with valve orifice

Warranty: 1 year warranty of replacement against any manufacturing defects.

Term and Condition:

1. Prices should be inclusive of all taxes, government levies, transportation, installation or any other charges should be specified separately & F.O.R.D.D. University, Nadiad.
2. Delivery: Within 15 weeks
3. Payment: 100% after satisfactory delivery / installation or through bank guarantee after delivery at D.D. University.
4. Validity: Quotation should be valid for 30 days after last date of submission.
5. Specify other term & conditions clearly with quotation.

Last date of receiving Tender is: 23/09/2019

The tender may please be super scribed to: Research & Development Center, DDU.

The tender may please be addressed to:
The Vice Chancellor
Dharmsinh Desai University,
P.O. Box No. 35, College Road,
Nadiad- 387001



(Dr. H. M. Desai)
Vice Chancellor

Dharmsinh Desai University

College Road, Nadiad - 387 001, India. Ph : 91 0268 2520502 Fax : 91 0268 2520501 Website : www.ddu.ac.in

NAAC ACCREDITED



Dharmsinh Desai University
R & D Center, Faculty of Technology, Nadiad.

Tender Notification

Date: 31/08/2019

Sealed tenders are invited for equipments as under,

Item 1: Photo Multiplier Tube (PMT)		Quantity
Specification:	• Gain of 10^5	01
	• Workable in the range 300 to 700 nm	
	• Small, compact and with housing	

Warranty: 1 year warranty of replacement against any manufacturing defects.

Term and Condition:

1. Prices should be inclusive of all taxes, government levies, transportation, installation or any other charges should be specified separately & F.O.R.D.D. University, Nadiad.
2. Delivery: Within 15 weeks
3. Payment: 100% after satisfactory delivery / installation or through bank guarantee after delivery at D.D. University.
4. Validity: Quotation should be valid for 30 days after last date of submission.
5. Specify other term & conditions clearly with quotation.

Last date of receiving Tender is: 23/09/2019


The tender may please be super scribed to: Research & Development Center, DDU.

The tender may please be addressed to:

The Vice Chancellor
Dharmsinh Desai University,
P.O. Box No. 35, College Road,
Nadiad- 387001

(Dr. H. M. Desai)

Vice Chancellor



Dharmsinh Desai University

College Road, Nadiad - 387 001, India. Ph : 91 0268 2520502 Fax : 91 0268 2520501 Website : www.ddu.ac.in

NAAC ACCREDITED